

PATHOZYME® PANCREATIC/GUT CANCER ANTIGEN 19-9 ^{Ref} OD277 Enzyme Immunoassay for the quantitative determination of Gastrointestinal CA19-9 in human serum. Store at 2°C to 8°C. DO NOT FREEZE. For in-vitro diagnostic use only.

INTRODUCTION

Pancreatic cancer accounts for approximately 2-3% of all cancers and the fourth most common cause of cancer deaths. The disease is more common in men than women with peak incidence occurring at age 60. CA 19-9 is the key marker for this disease.

CA 19-9 is a glycoprotein with a molecular weight of 1000000 Da. It is a derivative of the Le^a blood group antigen present on normal mesenteric epithelium. Patients that are Lewis a-/b- do not synthesise CA19-9. Recent reports indicate that the serum CA19-9 level is frequently elevated in the serum of subjects with various gastrointestinal malignancies, such as pancreatic, colorectal, gastric and hepatic carcinomas. Together with CEA, elevated CA19-9 is suggestive of gallbladder neoplasm in the setting of inflammatory gallbladder disease.

CA19-9 can also be elevated in non-malignant conditions such as chronic pancreatitis, cholangitis, cholecystitis and cirrhosis. For this reason the serum levels of CA19-9 should not be used in screening for pancreatic cancer but is of most use in the monitoring of the course of the disease and response to treatment.

INTENDED USE

PATHOZYME PANCREATIC/GUT CANCER ANTIGEN 19-9 is an Enzyme Immunoassay (EIA) for the quantitative determination of Gastrointestinal Cancer Antigen 19-9 in human serum.

For professional use only.

PRINCIPLE OF THE TEST

Specific monoclonal anti-Cancer Antigen 19-9 antibodies are coated on to microtitre wells. Test sera and Assay Buffer are added to the wells and incubated. The wells are then washed with water. Anti-Cancer Antigen 19-9 labelled with Horseradish Peroxidase enzyme (Conjugate) is added. If human Cancer Antigen 19-9 is present in the sample, it will combine with the antibody on the well and the enzyme Conjugate, resulting in the Cancer Antigen 19-9 molecules being sandwiched between the solid phase and the enzyme linked antibodies. After incubation, the wells are washed with distilled water to remove unbound labelled antibodies. On addition of the Substrate (TMB), a colour will develop only in those wells in which enzyme is present, indicating the presence of Cancer Antigen 19-9. The enzyme reaction is stopped by the addition of Stop Solution and the absorbance is then measured at 450nm. The concentration of Cancer Antigen 19-9 is directly proportional to the colour intensity of the test sample. This test has been calibrated against in house standards. There is no International standard for this test.

CONTENTS

^{Ref}
OD277



Microtitre Plate	12 x 8 wells x 1
Breakable wells coated with specific antibody contained in a resealable foil bag with a desiccant.	
BUF AS	11 ml
Sample Diluent: Phosphate based buffer containing stabilising proteins. Working strength. (Green)	
Cal A 0 U/ml	0.5 ml
Reference Standard: Cancer antigen 19-9 diluted in human serum. Ready to use. (Colourless)	
Cal B 25 U/ml	0.5 ml
Reference Standard: Cancer antigen 19-9 diluted in human serum. Ready to use. (Colourless)	
Cal C 75 U/ml	0.5 ml
Reference Standard: Cancer antigen 19-9 diluted in human serum. Ready to use. (Colourless)	
Cal D 150 U/ml	0.5 ml
Reference Standard: Cancer antigen 19-9 diluted in human serum. Ready to use. (Colourless)	
Cal E 300 U/ml	0.5 ml
Reference Standard: Cancer antigen 19-9 diluted in human serum. Ready to use. (Colourless)	
Cal F 600 U/ml	0.5 ml
Reference Standard: Cancer antigen 19-9 diluted in human serum. Ready to use. (Colourless)	
Conj 12 X	1 ml
Anti-Cancer Antigen 19-9 HRP Conjugate: Anti-Cancer Antigen 19-9 conjugated to HRP. (Yellow)	
DIL Conj	11 ml
Conjugate Diluent: Phosphate based buffer containing stabilising proteins. Working strength. (Red)	
Subs TMB	11 ml
Substrate Solution: 3,3',4,5'-Tetramethyl Benzidine in a citrate buffer. Ready to use. (Colourless)	
Soln Stop HCl 1M	11 ml
Stop Solution: Hydrochloric Acid diluted in purified water. Ready to use. (Colourless)	
Instruction Leaflet and EIA Data Recording Sheet	1 +1

MATERIAL REQUIRED BUT NOT PROVIDED

Micropipettes: 100µl, 200µl and 1000µl
 Disposable pipette tips
 Incubator: Temperature of 37°C +/- 1°C
 Absorbent paper
 Microplate reader fitted with a 450nm filter
 Graph paper
 Thoroughly clean laboratory glassware.

PRECAUTIONS

PATHOZYME PANCREATIC/GUT CANCER ANTIGEN 19-9 contains materials of human origin which have been tested and confirmed negative for HCV, HIV I and II antibodies and HBsAg by approved procedures at single donor level. Because no test can offer complete assurance that products derived from human source will not transmit infectious agents it is recommended that the reagents within this kit be handled with due care and attention during use and disposal. All reagents should, however, be treated as potential Biohazards in use and for disposal. Do not ingest

PATHOZYME PANCREATIC/GUT CANCER ANTIGEN 19-9 reagents do not contain dangerous substances as defined by current UK Chemicals (Hazardous Information and Packaging for Supply) regulations. All reagents should, however, be treated as potential biohazards in use and disposal. Final disposal must be in accordance with local legislation.

PATHOZYME PANCREATIC/GUT CANCER ANTIGEN 19-9 Stop Solution is dilute Hydrochloric Acid and is therefore corrosive. Handle with care. In case of contact, rinse thoroughly with water.

PATHOZYME PANCREATIC/GUT CANCER ANTIGEN 19-9 contain 1% Proclin™ 300* as a preservative which may be toxic if ingested. In case of contact, rinse thoroughly with running water and seek medical advice.

*Proclin™ 300 is a trade mark of ROHM & HAAS Limited.

STORAGE

Reagents must be stored at temperatures between 2°C to 8°C.

Expiry date is the last day of the month on the bottle and the kit label. The kit will perform within specification until the stated expiry date as determined from date of product manufacture and stated on kit and components. Do not use reagents after the expiry date.

Exposure of reagents to excessive temperatures should be avoided. Do not expose to direct sunlight.

DO NOT FREEZE ANY OF THE REAGENTS as this will cause irreversible damage.

SPECIMEN COLLECTION AND PREPARATION

Obtain a sample of venous blood from the patient and allow a clot to form and retract. Centrifuge clotted blood sample and collect clear serum. Fresh serum samples are required.

Do not use haemolysed, contaminated or lipaemic serum for testing as this will adversely affect the results.

Serum may be stored at 2°C to 8°C for up to 48 hours prior to testing. If longer storage is required, store at -20°C for up to 1 year. Thawed samples must be mixed prior to testing.

Do not use Sodium Azide as a preservative as this may inhibit the Peroxidase enzyme system.

Do not repeatedly freeze-thaw the specimens as this will cause false results.

REAGENT PREPARATION

All reagents should be brought to room temperature (20°C to 25°C) and mixed gently prior to use. Do not induce foaming.

Working Solution:
 Dilute the concentrated conjugate using 1 part conjugate to 11 parts conjugate diluent (1/12 dilution). 100µl is required per well. Prepare fresh working solution for each assay performed.

LIMITATIONS OF USE

The use of samples other than serum has not been validated in this test. There is no reuse protocol for this product.

When making an interpretation of the test it is strongly advised to take all clinical data into consideration. Diagnosis should not be made solely on the findings of one clinical assay.

ASSAY PROCEDURE

1. Bring all the kit components and the test serum to room temperature (20°C to 25°C) prior to the start of the assay.
2. One set of Standards should be run with each batch of test serum. Secure the desired number of coated wells in the holder. Record the position of the standards and the test serum on the EIA Data Recording Sheet provided.
3. Unused strips should be resealed in the foil bag containing the desiccant, using the resealing zip-lock before being replaced at 2°C to 8°C.
4. Dispense 10µl of Standards or test serum into the assigned wells.
5. Dispense 100µl of Assay Buffer into each well.
6. Thoroughly mix for 30 seconds. It is very important to have a complete mixing at this stage.
7. Incubate for 90 minutes at 37°C.
8. At the end of the incubation period, discard the contents of the wells by flicking plate contents into a Biohazard container. Then strike the wells sharply against absorbent paper. Ensure adequate disinfectant is contained in the Biohazard container.
9. Hand Washing: Fill the wells with a minimum of 300µl of distilled water per well. Flick plate contents into a Biohazard container. Then strike the wells sharply against absorbent paper. Wash the empty wells 5 times.
10. Strike the wells sharply onto absorbent paper or paper towel to remove all residual water droplets.
11. Machine Washing: Ensure that 300µl of distilled water is dispensed per well and that an appropriate disinfectant is added to the waste collection bottle. Wash the empty wells 5 times. After washing remove excess fluid by striking the wells sharply onto absorbent paper or paper towel to remove all residual water droplets.
12. Dispense 100µl working strength Conjugate solution into each well and mix gently for 10 seconds.
13. Incubate for 90 minutes at 37°C.
14. Wash plate as above.
15. Dispense 100µl of Substrate Solution into each well and gently mix for 10 seconds.
16. Incubate in the dark for 20 minutes at room temperature (20°C to 25°C).
17. Stop the reaction by adding 100µl Stop Solution to each well.
18. Gently mix for 30 seconds to ensure that the blue colour changes completely to a yellow colour.
19. Read the optical density immediately (no later than 10 minutes) using a microplate reader with a 450nm filter.

TROUBLESHOOTING

For use by operatives with at least a minimum of basic laboratory training.

Do not use damaged or contaminated kit components.

Use a separate disposable tip for each sample to prevent cross contamination.

Duplication of all standards and specimens, although not required, is recommended.

Specimens and standards should be run at the same time to keep testing conditions the same.

It is recommended that no more than 32 wells be used for each assay run if manual pipetting is used, since pipetting of all Standards and specimens should be completed within 3 minutes. A full plate of 96 wells may be used if automated pipetting is available.

Replace caps on all reagents immediately after use.

Avoid repeated pipetting from stock reagents as this is likely to cause contamination.

Do not mix reagents or antibody coated strips from different kits. When dispensing, care should be taken not to touch the surface of the well.

Do not allow reagent to run down the sides of the well. Prior to the start of the assay bring all reagents to room temperature (20°C to 25°C). Gently mix all reagents by gentle inversion or swirling.

Once an assay has been initiated, the wells should not be allowed to become dry during the assay.

The washing procedure is critical to the outcome of this test. Insufficient washing will result in poor precision and falsely elevated absorbance readings.

Do not contaminate the Substrate Solution as this will render the whole kit inoperative.

Check the precision and accuracy of the laboratory equipment used during the procedure to ensure reproducible results.

The unused strips should be resealed in the foil bag, containing the desiccant, using the resealing zip-lock before being replaced at 2°C to 8°C.

CALCULATION OF RESULTS

Calculate the mean absorbance value (A_{450}) for each set of Standards and specimens. Construct a standard curve by plotting the mean absorbance from each Standard against its concentration in U/ml on graph paper. Use the mean absorbance values for each specimen to determine the corresponding concentration of Pancreatic/Gut Cancer Antigen 19-9 in U/ml from the standard curve. If levels of controls or users known samples do not give expected results, test results must be considered invalid. If using a software package, choose a quadratic regression curve.

EXPECTED VALUES AND SENSITIVITY

The graph produced by the calibrators should be Hyperbolic in shape with the OD450 of the calibrators proportional to their concentration. The OD of calibrator A should be less than 0.5 and the OD of calibrator F greater than 1.2 for the assay results to be valid. Healthy men and women have normal expected Cancer Antigen 19-9 concentrations below 35 U/ml. The minimum detectable concentration of Cancer Antigen 19-9 by **PATHOZYME PANCREATIC/GUT CANCER ANTIGEN 19-9** is estimated to be 10U/ml.

EVALUATION DATA

Calibrated to major competitors and in house standards.

The co-efficient of variation of **PATHOZYME PANCREATIC/GUT CANCER ANTIGEN 19-9** is less than or equal to 10%.

In an evaluation between the Omega Pathozyme CA 19-9 kit and the Abbott IMx Kit for samples with levels between 33.4 U/ml and 297.6 U/ml the following data was generated.

Number of Samples	44
Correlation Co-efficient	0.9738
Slope	0.94
Intercept	6.77
Omega Mean	100.38 U/ml
Abbott Mean	99.33 U/ml

These kits were shown to give good correlation.

REFERENCES

1. Glenn, J., Steinberg, W.M., Kurtzman, S.H., et al. Evaluation of the utility of a radioimmunoassay for serum Cancer Antigen 19-9 level in patients before and after treatment of carcinoma of the pancreas. *J. Clin. Oncol.* 1988;6:462-8.
2. Hayakawa, T., Kondo, T., Shibata, T. et al. Sensitive serum markers for detecting pancreatic cancer. *Cancer* 1988;61:1827-31.
3. Koprowski, H., Herly, Steplewski, Z., et al. Specific antigen in serum of patients with colon carcinoma. *Science* 1981;212:53-55.
4. Malesci, A., Tommasini, M.A., Bonato, C. et al. Determination of Cancer Antigen 19-9 antigen in serum and pancreatic juice for differential diagnosis of pancreatic adenocarcinoma from chronic pancreatitis. *Gastroenterology* 1987;92:60-7.
5. Safi, F., Roscher, R., Bittner, R., et al. High sensitivity and specificity of CA 19-9 for pancreatic carcinoma in comparison to chronic pancreatitis. Serological and immunohistochemical findings. *Pancreas* 1987;2:398-403.
6. Steinberg, W. The clinical utility of CA 19-9 tumour associated antigen *American J. of Gastroenterology* 1990;85:350-355.
7. Steinberg, W.M., Gelfand, R., Anderson, K.K., et al. Comparison of the sensitivity and specificity of the CA 19-9 and Carcinoembryonic antigen assays in detecting cancer of the pancreas. *Gastroenterology* 1986;90:343-9.
8. Takasaki, H., Uchida, E., Tempero, M.A., et al. Correlative study on expression of CA 19-9 and DU-Pan-2 in tumour tissue and in serum of pancreatic cancer patients. *Cancer Res.* 1988;48:1435-8.
9. Tetsuta, M., Yamamura, H., Iishi, H., et al. Values of Cancer Antigen 19-9 in the serum, pure pancreatic juice and aspirated pancreatic material in the diagnosis of malignant pancreatic tumour. *Cancer* 1985;56:2669-73.
10. Wang, T.H., In Chen, J.W. et al. Non-Invasive diagnosis of advanced pancreatic cancer by real-time ultra sonography, Carcinoembryonic antigen and carbohydrate antigen 19-9. *Pancreas* 1986;1:1219-23.
11. Strom B.L., Maislin, G., West, S. I. et al. Serum CEA and Cancer Antigen 19-9: potential future diagnostic or screening tests for gallbladder cancer? *Int. J. Cancer* 1990;45:821.

QUICK REFERENCE TEST PROCEDURE

1. Dispense 10µl of samples or Standards and 100µl of Assay Buffer into each well and gently mix for 30 seconds.
2. Incubate for 90 minutes at 37°C.
3. Discard the well contents and wash 5 times with distilled water.
4. Dispense 100µl of working strength Conjugate solution into each well and gently mix for 10 seconds.
5. Incubate for 90 minutes at 37°C.
6. Discard the well contents and wash 5 times with distilled water.
7. Add 100µl of Substrate Solution to each well. Gently mix for 10 seconds.
8. Incubate in the dark for 20 minutes at room temperature (20°C to 25°C).
9. Add 100µl of Stop Solution to each well and gently shake for 30 seconds.
10. Read Optical Densities immediately (no later than 10 minutes) using a microplate reader with a 450nm filter.

8077 ISSUE 8 Revised February 2010.

©Omega Diagnostics Ltd 2010.

